



PowerBeat M Series Automated External Defibrillator User Manual



Before Use

Thank you for purchasing the PowerBeat M Series Automated External Defibrillator.

Please read this manual carefully before use. Please keep the device for easy access after use.

Version: 1.0

Date: 2023.09.11

Manufacturer



ViVest Medical Technology Co., Ltd.

Unit 205, 206, 207, B1 Building, SIP BioBay, No.218, Xinghu Street, Suzhou Industrial Park, Suzhou Area, Suzhou Free Trade Pilot Zone, Jiangsu, China.

TEL: +86-0512-65730937

FAX: +86-0512-65730937

Email: service@vivest.cn

Authorized EC Representative



Shanghai International Holding Corp. GmbH (Europe)

Add: Eiffestrasse 80, 20537 Hamburg, Germany

Tel: +49-40-2513175

Fax: +49-40-255726

E-mail: shholding@hotmail.com

Copyright and Statement

This manual applies to the PowerBeat M Series Automated External Defibrillator.

The copyright of this manual is owned by ViVest Medical Technology Co., Ltd. (Hereinafter called "VIVEST"). No organization or individual may reproduce this manual or any of its contents in any form without the company's permission.

The company does not assume any responsibility for any injury caused by failing to follow the instructions, precautions, warnings, or usage instructions in this manual.

The copyright of the software in this product is owned by VIVEST. This software is protected by copyright laws and international treaty provisions that apply throughout the world. Without the permission of the company, any organization or individual may not copy, decompile, reverse engineer, or disassemble this software into a form that people can understand. VIVEST reserves the right to own the software.

For information concerning any of our products, please contact VIVEST.

Conventions

This manual uses the following conventions:

Within text, the name and labels for physical buttons and softkeys appear in boldface type (for example, "Press the **Shock** button").

This manual uses italics for audible prompts (for example, "*Don't touch patient, Analyzing*").

Content

I	General Introduction	1
	Indication of Use	1
	Contraindication of Use	1
	Intended Use	1
	Intended User	1
	Product Features	1
	Product Limitations	2
II	Safety Precautions	3
	Classification of Warning Messages	3
	Precaution Information	3
	Placement of the Device	5
III	Installation and Preparation	7
	Unpacking	7
	Control Panel	7
	Install the Battery	10
	Remove the Battery	10
	Charge	11
	Preconnect the Pads	11
	Self-test System	12
IV	Use Automated External Defibrillator	14
	Brief operation steps	14
	Turn on the Device	14
	Apply Pads	15
	To apply pads:	15
	Cardiac Rhythm Analysis	17
	Shock Advised	18
	No Shock Advised	18
	Perform CPR	19
	Operation After Use	20
	Pediatric Treatment	20
V	Maintenance and Troubleshooting	21
	Regular Maintenance	21
	Check Pads	21
	Check Status Indicator	22
	Check Integrity and Cleanliness	22

Check Battery	22
Cleaning	22
Transport	22
Disposal	22
Troubleshooting	23
VI Cybersecurity	25
Network Environment	25
Data Interface	25
User Access	25
Data Exchange Method	26
Bluetooth Transmission	26
4G Transmission	26
Device Security Software	26
Cyber Security Update	26
AED Data Storage	26
ViBest-AED Data Management System	27
Appendix 1 Accessories	A
Appendix 2 Symbols	B
Appendix 3 Glossaries	F
Appendix 4 Specifications	H
Appendix 5 Defibrillation Waveform	L
Appendix 6 ECG Analysis System	N
Appendix 7 EMC Guide	P

I General Introduction

Indication of Use

PowerBeat M Series Automated External Defibrillator should be applied only when the patient has the following symptoms at the same time:

- Unconsciousness
- Not breathing or abnormal breathing
- Unresponsiveness

Contraindication of Use

PowerBeat M Series Automated External Defibrillator can not be use if the patient:

- Is conscious,
- Is breathing, or
- Has a detectable pulse or other sign of circulation

Intended Use

The device is intended to be used for external defibrillation in patients with suspected cardiac arrest who are unresponsive, non-breathing, and pulse-free. Adult mode is suitable for patients who are not younger than 8 years old, and child mode is suitable for patients under 8 years old or weighing less than 25kg.

This product is intended for use in public or medical settings by qualified personnel trained in the use of CPR and AED, or by medical personnel who have received basic life support/advanced life support Use under the guidance of the central dispatcher.

Intended User

The device is intended to be used by rescuers and emergency care personnel who have completed training and certification requirements applicable to the use of a defibrillator where the operator controls delivery of shocks to the patient.

Service personnel must be authorized by the manufacturer, trained and qualified, and have a full knowledge and understanding of the contents of this manual.

Product Features

The PowerBeat M Series Automated External Defibrillator is divided into 2 models, PowerBeat M1 and PowerBeat M3 (Hereinafter referred to 'the device'). The 2 models have similar functions except below the following:

Model	PowerBeat M1	PowerBeat M3
Function Button	✓	✗
Child Button	✗	✓

PowerBeat M1 supports 4 operating modes: Standby mode, Rescue mode, Continuous VF/VT recognition mode, and Maintenance mode.

PowerBeat M3 supports 3 operating modes: Standby mode, Rescue mode and Maintenance mode

The mainly used functions and features of the device are shown as follow:

Function	Model	PowerBeat M1	PowerBeat M3
Voice and Light Guidance		✓	✓
Heart Rhythm Analysis		✓	✓
Defibrillation		✓	✓
Self-test System		✓	✓
Continuous VF/VT Recognition Function		✓	✘
Volume Control		✓	✓
Child Mode (under 8 years old or weighing less than 25kg)		✘	✓
Record		✓	✓
Data Transmission		✓	✓

Product Limitations



The device is an infrequently-used device, it has certain limitations as outlined below:

- Routine maintenance is needed to ensure the device is ready for use. See Chapter 5 for details.
- This device is not intended for use in patients with implanted and activated ICDS.
- This device should not be used in an MRI environment.

II Safety Precautions

Classification of Warning Messages

Warning Messages are generally divided into 3 categories, as described below:

 Danger	Warning statements alert you to conditions or actions that can result in personal injury or death.
 Warning	Indicates potential risks or risks caused by unsafe operations, which could result in personal injury or property damage if not avoided.
Caution	Used to emphasize instructions or reminders so that users can operate this device safely.

Precaution Information

Danger

- 1) The device comes out with high voltage during defibrillation and may cause severe personal injury or even death. Therefore, defibrillation should be performed by a professionally trained layperson.
 - 2) Components replacement can only be performed by the manufacturer. Other personnel must not attempt to open the shell to repair the device and replace components. Otherwise, there is a risk of electric shock.
 - 3) Do not disassemble or modify the device without the manufacturer's authorization, or it may lead to personal injury or even death.
 - 4) Nobody should touch the patient during defibrillation, and other MEDICAL EQUIPMENT which has no DEFIBRILLATION-PROOF insulation should be disconnected from the PATIENT during defibrillation.
 - 5) During defibrillation, move away from the patient and remove all metal equipment connected to the patient, failure to do so may result in an electric shock.
 - 6) There may be a danger of electric shock or personal injury if the defibrillation energy is not released normally.
 - 7) To prevent fire or explosion, do not use the device in an environment with flammable gases or concentrated oxygen.
 - 8) Be sure not to charge the disposable battery. Charging the disposable battery may cause a fire or explosion. When charging the rechargeable battery, be sure to use a dedicated charger.
 - 9) Do not burn or incinerate the battery. Burning or incinerating the battery may cause a fire or explosion.
 - 10) Do not perform maintenance on the device during use.
 - 11) Do not remove the battery when the device enters rescue mode or when the device is placed in public places.
 - 12) Improper operation may cause the battery to heat up, catch fire, or explode. Please read the warnings carefully before using rechargeable batteries.
-

-
- 13) If you find that the battery leaks or gives off a bad smell, stay away immediately. If the electrolyte leaks into the skin or clothing, wash it immediately with water. If the electrolyte seeps out and gets into your eyes, do not rub your eyes, wash them immediately with clean water and see a doctor.
 - 14) In order to ensure the expected life of the battery, if the rechargeable battery is not used for a long time, please charge the rechargeable battery at least every three months.
 - 15) When the battery reaches its expiration date, or when the battery is found to be smelly, deformed, discolored or distorted, the battery should be stopped from being used and disposed of in accordance with local regulations.
-

 **Warning**

- 1) The defibrillation can only be performed by those technicians who are professionally trained and familiar with the operation of the device.
 - 2) Ensure the device is carefully placed to avoid damage to the pads or device, or injury to the patient or operator during use.
 - 3) To prevent the device from falling or dropping, the device should be placed and affixed in a position that prevents it from falling or dropping. If the device falls or is dropped, it must be checked immediately for any damage.
 - 4) Do not use expired or dry pads as they will not completely adhere to the skin.
 - 5) Do not repeatedly or rapidly charge and discharge the device except as necessary during the rescue. If the device test requires repeatedly internal discharges, wait at least one minute after every 3 discharges.
 - 6) Do not connect the pads to other pads or metal objects in contact with the patient, for the conductive gel coating on the pads may stick to other objects. It is recommended to keep a distance of at least 5cm. Defibrillation with an insufficient conductive gel may cause a skin burn.
 - 7) Before defibrillation, shave body hair from the patient's chest if necessary. Excessive body hair may cause skin burns.
 - 8) Do not wipe the patient's skin with alcohol. Alcohol will dry the skin and cause skin burns.
 - 9) If the patient has an implanted pacemaker, do not place the pads directly near the implanted pacemaker.
 - 10) Do not use the device if the device has been soaked with liquid or lots of water on the surface. The conductive part of the device must not be in contact with other conductive parts (including the ground).
 - 11) To avoid accidental electric shock, no functional examination may be performed when the device is connected to the patient.
 - 12) Do not use alcohol or other solutions to soak or clean the pads. This may damage the pads and cause the device to malfunction.
 - 13) Moving or carrying the patient during heart rhythm analysis can cause diagnostic delays or errors.
 - 14) Pads should adhere to a flat skin surface instead of the folded skin surface. If the pads are not placed in the right place, it will affect the heart rhythm analysis and may cause wrongful defibrillation.
 - 15) When using the device, the operator must keep the PATIENT's body (such as exposed skin or head and limbs) away from touching conductive fluids (such as gel, blood, or saline) and metal objects (such as a bed frame or a stretcher), to prevent alternate pathways for the defibrillation
-

current.

- 16) Do not place the device near an apparatus that emits strong Radio-Frequency(RF) signals. Radiofrequency emissions cause incorrect analysis of the heart rhythms.
 - 17) Use pads, batteries, and other accessories provided by VIVEST only. The company is not responsible for any problems caused by the use of components not approved by VIVEST.
 - 18) The device cannot work if the battery is flat and/or uninstalled. Replace the battery immediately if the device is detecting low battery power or the battery is overdue for replacement.
 - 19) If the device is taken out from the highest storage temperature or the lowest storage temperature and put into use immediately, the performance of the device may differ from expectations. The device must not be stored or used outside of the environmental limits specified in this manual.
 - 20) Improper operation may cause runtime errors. Please follow this manual strictly.
 - 21) Only the service personnel should configure the device to use Bluetooth. The use of Bluetooth will not result in any risk to the device or its use.
 - 22) When the status indicator of the device is found to be off, the service personnel should try to replace the battery to return the device to normal.
 - 23) Be sure not to touch the patient during defibrillation, or it may cause a risk of electric shock.
-

Caution:

- 1) If any damage occurs to the device, please contact the manufacturer for repair.
- 2) Please pay attention to all caution and warning signs on the device and accessories.
- 3) If the device is stored, transported, or used outside the limited environment range, the performance in this user manual may not be achieved.
- 4) The device can be expected to be operated under 50°C, but it is recommended to use under 40°C to avoid burning patients.
- 5) It is recommended to provide at least one extra battery for each device available in a public place.

Placement of the Device

The device can be fixed or carried according to user needs:

When fixed, the device should be placed around emergency equipment (such as fire extinguishers, first aid kits, etc.) at a suitable temperature, and keep away from moisture and dust. To ensure correct placement of the device:

- 1) The ambient temperature at which the device is placed should be between 0°C and 50°C (long term placement). Severe fluctuation of the ambient temperature may significantly shorten the battery's service life and affect the performance of the pads.
- 2) It should be stored in a dry place with a relative humidity of 5% to 95%.
- 3) It should be stored away from any direct sunlight. Long-time exposure to direct sunlight will accelerate the aging of the device.
- 4) To avoid the blockage of the speaker, the device should not be placed in an environment with lint or dust.
- 5) Do not place the device near a strong magnetic field.

When carrying, the device shall be equipped with specialized portable package, and follow the below contents:

1) The environment temperature should be at 0°C to 50°C, relative humidity environment should be in 0% to 95%, temperature or humidity out of scope may shorten the service life and performance of AED.

2) When carrying, do not close to strong magnetic field.

III Installation and Preparation

This chapter mainly introduces the components and appearance structure of the device, the functions of the buttons and indicators of the control panel, and the installation of key components.

Unpacking

To ensure the integrity of the device, carefully take out all components from the packaging case and follow the steps below to check the device:

- 1) Check the intact of the device shell.
- 2) Check whether the pads pre-connect to the device.
- 3) Check the seal and expiration date of the pads.
- 4) Check the expiration date of the battery.

Control Panel

The PowerBeat M1 control panel is shown below:




Figure III- 1 PowerBeat M1 control panel

The graphic description:

Name	Description
A: Power Button	In standby mode, press the power button to enter rescue mode. In rescue mode, press the power button for at least 2 seconds to return to standby mode.
B: Status Indicator	Indicates the current status of the device: <ul style="list-style-type: none"> ➤ Green light flash shows that the device is in standby mode and ready to use. ➤ Red light flash shows that the device fails to pass the self-test and needs maintenance. ➤ Green lights steady shows that the device is in use. ➤ Light off shows that the battery is not installed or the device is in abnormal situation.

Name	Description
C: Pads Indicator	Indicates the status of pads connection: <ul style="list-style-type: none"> ➤ Light off shows that the pads are connected normally. ➤ Red light flash shows that the pads are not connected or connected abnormally.
D: Heart Rhythm Analysis Indicator	Indicates the status of heart rhythm analysis: <ul style="list-style-type: none"> ➤ Green light flash shows that the device is analyzing heart rhythm. ➤ Light off shows that the device is not in analyzing stage.
E: Shock Button	Press the shock button to deliver defibrillation energy after the charge is finished.
F: Shock Indicator	Orange light flash to guide operators to press shock button when charge is finished.
G: CPR Indicator	Indicates the device is in the CPR stage.
H: Function Button	1) Adjust the volume. Press the function button less than 1 second to switch the maximum volume and minimum volume to each other. 2) Switch the mode. Press the function button for at least 3 seconds to switch rescue mode and continuous VF/VT recognition mode to each other.
I: Volume Indicator	Blue lights steady indicates that the volume of the device is minimum.
J: Continuous VF/VT Recognition mode Indicator	White lights on indicates that the device is in continuous VF/VT recognition mode.
K: Speaker	Sends voice prompts and beep sounds.
L: Battery	Supply power to the device.
M: Pads Cable Connector	Pads connect to the host through the connector.
N: Pads Storage Box	To store the pads.
O: Microphone	Used optionally to record audio during rescue.

The PowerBeat M3 control panel is shown below:




Figure III-2 PowerBeat M3 control panel

The graphic description:

Name	Description
A: Power Button	<p>In standby mode, press the power button to enter rescue mode.</p> <p>In rescue mode, press the power button for at least 2 seconds to return to standby mode.</p>
B: Status Indicator	<p>Indicates the current status of the device:</p> <ul style="list-style-type: none"> ➤ Green light flash shows that the device is in standby mode and ready to use. ➤ Red light flash shows that the device fails to pass the self-test and needs maintenance. ➤ Green lights steady shows that the device is in use. ➤ Light off shows that the battery is not installed or the device is in abnormal situation.
C: Pads Indicator	<p>Indicates the status of pads connection:</p> <ul style="list-style-type: none"> ➤ Light off shows that the pads are connected normally. ➤ Red light flash shows that the pads are not connected or connected abnormally.
D: Heart Rhythm Analysis Indicator	<p>Indicates the status of heart rhythm analysis:</p> <ul style="list-style-type: none"> ➤ Green light flash shows that the device is analyzing heart rhythm. ➤ Light off shows that the device is not in analyzing stage.
E: Shock Button	<p>Press the shock button to deliver defibrillation energy after the charge is finished.</p>
F: Shock Indicator	<p>Orange light flash to guide operators to press shock button when charge is finished.</p>
G: CPR Indicator	<p>Indicates the device is in the CPR stage.</p>
H: Child Button	<p>If you need to enter the child mode, press this button, and the device will voice into the child mode. Please hold down the child button for 3 seconds, and hold down this button for 3 seconds to switch the device to the child mode.</p> <p>(To return to the adult mode, the device should shut down and</p>

Name	Description
	restart).
I: Child mode Indicator	Flash blue indicates long press the child button to enter child mode. Blue lights on indicates that the child energy is active.
J: Speaker	Send voice prompts.
K: Battery	Supply power to the device.
L: Pads Cable Connector	Pads connect to the host through the connector.
M: Pads Storage Box	To store the pads.
N: Microphone	Used optionally to record audio during an incident.

Install the Battery

The device accepts the following battery:

- Disposable Battery (LiMnO₂)
- Rechargeable Battery (Li-ion)

To install the battery:

- 1) Push the battery tail into the battery compartment.
- 2) Push the battery to the bottom of the battery compartment.
- 3) Finally, check whether the buckle of the battery is fully inserted into the battery slot.




Figure III-3 Install the battery

After the battery is installed, the device will run self-test automatically. Refer to Chapter 3.6 for details.

Remove the Battery

When the 'Low Battery' prompt is indicated, please replace the battery immediately. After removing the battery, please wait for 30 seconds before installing new battery.

To remove the battery:

- 1) Make sure the machine is in standby mode, if the device is in rescue mode, press the **Power** button for at least 2 seconds to enter standby mode.

- 2) Press the battery buckle.
- 3) Immediately remove the battery.




Figure III-4 Remove the battery

Charge

The rechargeable battery is equipped with a dedicated battery charger that can hold two rechargeable batteries for charging at the same time.

- The battery charger is connected to the power supply and the battery is not inserted: The indicator flashes green slowly in the form of a breathing light.
- Battery charger failure: The indicator flashes red.
- Charging: The indicator light flashes green.
- Charge finished: The indicator lights green.

The battery charger shows below:



Figure III-5 Battery charger

Preconnect the Pads

The pads are pre-connected when the device leaves the factory, but it is still necessary to check whether the pads connector is inserted properly before use.

Before using, please check whether the pads package is in good condition and whether expired. If the package is damaged or the electrode is expired, please contact the dealer or manufacturer in time for replacement.

When inserting the pads connector into the socket, be sure to insert tightly.




Figure III-6 Connect pads to the host

Warning

- 1) Never use a damaged, wrinkled, and folded pads, or it may cause dangerous such as leakage current and unwanted burns on skin.
- 2) Don't reuse the disposable pads. Repeated use may cause performance degradation or cross infection.

Self-test System

The device performs the following self tests to verify unit integrity and its readiness for emergency use:

- User Self-test
- Battery Installation Self -test
- Power On Self-test
- Periodic Self-test
- Quick status check

User self-test

Service personnel authorized by manufacturer can run user self-test if necessary.

Battery installation self-test

The device performs a self-test whenever batteries are installed. Then the device enters into the standby mode at the end of the battery installation self-test.

- The status indicator blinks green in the standby mode shows that all the tests passed.
- Otherwise, the status indicator flash red, with 5 times of Beep sound, to warning the operator or service personnel.

Power on self-test

The device performs a self test whenever the unit is turned on, this will notify the operator of any failures identified during the self-test.

If detected...	Voice Prompt
Pads expired	<i>Pads Overdue</i>
Pads connector not detected	<i>Plug in Pads Connector</i>
Battery expired	<i>Battery Overdue</i>
Device failure	<i>Equipment Fail</i>

Battery is low	<i>Low Battery</i>
Battery is about to run out	<i>Low Battery, replace battery immediately</i>

Periodic self-test

Periodic self-test will be carried out daily, weekly, monthly, and quarterly. The default self-test configuration is weekly self-test, monthly self-test, and quarterly self-test. Service personnel can set daily self-test as required. The default period self-test time is 3 am.

Caution:

- 1) Only in the standby mode when the battery is installed, the device will automatically detect at a preset time.
- 2) If the device self-test result is normal, the operator can start rescue immediately.
- 3) If the device self-test result is abnormal, contact the maintenance personnel or the manufacturer for repair.

Quick status check

In standby mode, the device supports a quick status check by pressing the **Function** button of PowerBeat M1 or the **Child** button of PowerBeat M3 to check whether the device is capable of emergency use.

- If the check result is normal, the status indicator blinks green once.
- If the check result is abnormal, the status indicator flashes red once, and the device emits three beacons.

IV Use Automated External Defibrillator

This chapter mainly introduces usage of the device. There are voice and indicator prompt during rescue.


Brief operation steps

Rescue Mode:

1. Press the **Power** button to turn on the device.
2. Remove any clothing from the patient’s chest and take out pads.
3. Attach the pads to the patient according to the figure instruction.
4. Operate according to the voice prompt:
 - If a shockable heart rhythm is detected, press the **Shock** button to defibrillate.
 - If a non-shockable heart rhythm is detected, the device will enter CPR stage.
5. Perform CPR to patient.

Continuous VF/VT Recognition mode: (PowerBeat M1 Only)

1. When the device is in the CPR stage and the result of the last heart rhythm analysis is No Shock Advice, while the patient remains conscious (responsive, breathing and pulse), long press the **Function** button for > 3 seconds to enter the continuous VF/VT recognition mode
2. During continuous VF/VT recognition mode:
 - If a shockable heart rhythm is detected, the device will switch to rescue mode and re-analyzes the rhythm. If the analysis result is ‘shock advised’, the device will charge to preset energy and defibrillate, and then enters CPR stage.
 - If a non-shockable heart rhythm is detected, the device will re-analyze the rhythm.
 - Press the **Function** button for at least 3 seconds, the device will switch to rescue mode.

 Warning	<p>1) The use of this continuous VF/VT recognition pattern must be confirmed by a paramedic trained in its use in order to provide a non-diagnostic indication of shockable and non-shockable heart rhythm for patients who are responding to treatment and breathing normally, but may have discomfort in the cardiac area.</p> <p>2) In the continuous VF/VT recognition mode, the device continuously identifies the patient's heart rhythm to evaluate whether the patient has a shockable or non-shockable heart rhythm. In this mode, someone needs to supervise the patient so that if the patient's heart rhythm or health condition changes, the ambulance crew can effectively care for the patient.</p>
--	--

Turn on the Device

Press the power button to turn on the device.

Step	Action	Voice Prompt	Indicator
1	Press the Power button (Figure IV-1)	<i>Call for help</i>	Status indicator light stays green during all the rescue process.




Figure IV- 1 Press the power button

Apply Pads

To apply pads:

Step	Action	Voice Prompt	Indicator
1	Remove patient's upper body clothing. If necessary, remove the patient's chest hair and dry the patient's skin to ensure good contact between the electrode and the patient's skin. (Figure IV-2)	<i>Remove clothing</i>	Pads Indicator flash red till the pads are applied properly.
2	Take out the electrode bag from the pads storage box at the bottom of the device, then tear open the package and take out the pads. (Figure IV-3)	<i>Remove Pads Package from Back of AED</i> <i>Tear open package, take out the pads</i>	
3	Remove the liner of the pads and apply the pads on the exposed skin of the patient according to the diagram. The position should be consistent with the diagram on the pads. (Figure IV-4, Figure IV-5)	<i>Remove Liner from Pads</i> <i>Apply Pads to Patient's Chest</i>	




Figure IV- 2 Patient Preparations




Figure IV-3 Open pads storage box




Figure IV-4 Remove pads liner

When attaching the pads, firmly adhere to one side at first and then press it smoothly to the other side, make it completely attached to the skin to avoid air bubbles. Do this way for both two pads.

Adults and children (PowerBeat M3 only) share the same pads, but it is attached to different parts of the patient's body:

➤ Attachment position of electrode on child body:

Place one electrode in the middle of the chest between the nipples and the other in the middle of the back (anterior-posterior).

➤ Attachment position of electrode on adult body:

One of the electrodes was applied to the patient's right chest and below the clavicle as illustrated, and the other electrode was applied to the patient's left chest and above the rib in the maxillary line.

Those two positions are shown in the figure below:





Figure IV-5 pads position on the patient

 Warning	<p>1) The pads must be placed flat on the patient's skin. Not doing so may lead to an incorrect heart rhythm analysis and wrongful defibrillation.</p> <p>2) Leaving bubbles between the pads and the patient's skin when attaching the pads may result in burns.</p> <p>3) Make sure the pads have a good contact with the patient's body, a poor contact might cause skin burns.</p>
--	--

Cardiac Rhythm Analysis

While the pads are applied:

Step	Action	Voice Prompt	Indicator
1	The device will analyze the patient's heart rhythm as long as the pads are attached to the patient correctly.(Figure IV-6)	<i>Don't Touch Patient, Analyzing Heart Rhythm</i>	Heart rhythm analysis indicator flash green.
2	If the pads are not attached properly, the analysis will be interrupted.	<i>Poor Pads Contact, Check Pads</i>	Pads indicator flash red.
3	If signal interference is detected...	<i>Eliminate signal interference</i>	/
4	If motion interference is detected...	<i>Keep Patient Still</i>	/





Figure IV-6 Cardiac Rhythm Analysis

At this point, nobody is allowed to touch the patient or pads. Keep the distance between the patient and the device about 10cm to 90cm, and the distance between the operator and the patient should be less than the distance of one arm.

 Warning	Do not touch or shake the patient in the process of cardiac rhythm analysis, otherwise it will affect the result!
--	---

Shock Advised

While a shockable rehythm is detected:

Step	Action	Voice Prompt	Indicator
1	If a shockable heart rhythm is detected...	<i>Shock Advised</i> <i>Do not Touch Patient,</i> <i>Charging</i> <i>Press Flashing Shock Button</i>	Shock indicator flash orange.
2	Press the Shock button(Figure IV-7)	<i>Shock Delivered</i> <i>Begin CPR</i>	Shock indicator turned off and CPR indicator flash green.
3	If the Shock button is not pressed within 30 seconds	<i>Shock Button not Pressed,</i> <i>Shock Cancelled</i>	Shock indicator turned off.
4	If the patient's heart rhythm turns into a non-shockable rhythm within 30 seconds	<i>Heart Rhythm Changed,</i> <i>Shock Cancelled</i>	Shock indicator turned off.





Figure IV-7 Press Flashing shock button

The device will detect the pads connection continually during charge process, it will stop charge and prompt the operator to check the pads connection if poor contact is detected.

During the whole rescue process, please pay attention to the patient. If the patient wakes up, shut the device down immediately.

 Warning	Nobody should touch the patient during defibrillation!
--	--

No Shock Advised

While a non-shockable rhythm is detected:

Step	Action	Voice Prompt	Indicator
1	If a non-shockable heart rhythm is detected...	<i>No Shock Advised</i> <i>Begin CPR</i>	CPR indicator flash green

Perform CPR

After the defibrillation:

Step	Action	Voice Prompt	Indicator
1	Hands crossed and placed on the patient's chest, and follow the beat sound for chest compressions.(Figure IV-8)	<i>Beep...Beep...Beep...</i>	CPR indicator flash green
2	After 30 times of compress, give two breaths.	<i>Breathe...Breathe</i>	CPR indicator flash green
3	After 2 minutes of CPR...	<i>Stop CPR</i> <i>Don't Touch Patient,</i> <i>Analyzing Heart Rhythm</i>	CPR indicator turned off and heart rhythm analyze indicator flash green.




Figure IV-8 Start CPR

Adult mode contains 2 kinds of compression-to-ventilation ratio, while child mode contains 3 kinds of compression-to-ventilation ratio. Service personnel can modify the compression-to-ventilation ratio using the configuration software:

Mode	Compression-to-ventilation ratio
Adult mode	<ul style="list-style-type: none"> ✓ 30:2 (Default) • Only compression
Child mode	<ul style="list-style-type: none"> • 30:2 ✓ 15:2 (Default) • Only compression

When CPR is over, the device will reanalyze the patient's heart rhythm.

Operation After Use

After using the device, perform the following steps:

- 1) Press the **Power** button for at least 2 seconds to enter the standby mode
- 2) Check the appearance of the device for damage or dirt. If technical support for damage is needed, please contact the manufacturer. If it is dirty, please clean it up, refer to Chapter 5 for details.
- 3) Dispose of disposable defibrillation electrodes in accordance with local regulations after use, and take out the new defibrillation electrodes (check expiration date), put it into the electrode box of the device.
- 4) In order to ensure that the device has sufficient power during the next usage, please check the battery power. If the battery power is low, please replace the battery (or charge the battery).

Pediatric Treatment

For patients less than 25kg (55 lb.) or under 8 years old, most of the patients who have cardiac arrests are not from cardiac problems. The rescue steps shown below:

1. Begin CPR, ask someone to call the emergency center and the nearest AED, and perform CPR till get AED.
2. After obtaining AED, turn it on and press the child button to enter the child mode, the child indicator light starts to flash, the device will issue a voice prompt "*To enter the child mode, press the child button for 3 seconds*", press and hold the child button for 3 seconds, the device will send out The voice prompts "*Child mode*". At this time, the device switches to child Mode. The child mode indicator light is always blue and the defibrillation energy decreases from 150J to 50J.
3. Remove the patient's upper body clothing, expose the front chest and back, and place one electrode in the middle of the chest between the nipples and the other in the middle of the back.
4. Follow the voice prompts. If an electric shock is recommended, press the **Shock** button. Otherwise, the device will directly enter the CPR stage.

Caution:

Do not miss the best first aid time just to determine the patient's age. If the patient's age cannot be determined, treat the patient in an adult mode.

V Maintenance and Troubleshooting

This chapter describes the regular maintenance, transport, disposal, and troubleshooting of the device. Some of those operations should be guided by the authorized service personnel.

Regular Maintenance

The expected service life of the device is 10 years. In order to ensure the reliability of the device, service personnel should carry out routine maintenance and inspection of the device during the service period. If the machine is more than 5 years old, the frequency of routine maintenance and inspection should be increased appropriately.

The device minimizes required maintenance by using extensive self-tests to simplify the maintenance process. The device will monitor its essential performance automatically during use and run periodic self-test automatically in standby mode. Refer to Chapter 3.5 for details.

By visually checking the status indicator every day, the service personnel can know that whether the device has passed the self-test within the last 24 hours and confirm whether the device is ready for use. To calibrate the impedance and check the accuracy of discharge energy, please contact the manufacturer. By connecting to the device management system, you can remotely manage the device, reducing onsite maintenance. All maintenance work performed must comply with local regulations.

Daily maintenance	Monthly maintenance	After Rescue maintenance	Maintenance Content
✓	✓	✓	Check the status indicator
✓	✓	✓	Check device and accessories
		✓	Replace the pads
		✓	Check battery power and expiration date
		✓	Manual self-test

Warning

The device has **NO** user-serviceable components. ALL Components of the device can only be replaced or renewed by the manufacturer. If any problem occurs, contact the maintenance personnel or the manufacturer.

Check Pads

The defibrillation pads of the device are disposable. If the pads has been used or the package is damaged, contact local distributor or manufacturer to replace in time.

Please check the date on the package to confirm whether it is expired. Expired electrodes should be disposed of according to local regulations.

In addition, the device can detect the expiration date of the pads through self-test. If the pads expired, the status indicator flashes red in standby mode.

- Check whether the pads cable is damaged, if damaged, please replace the pads immediately.
- Check whether the pads connector has been inserted. If not, insert it into the connector socket.

Check Status Indicator

The device standby status indicator is located at the top center of the panel, which indicates the status of device.

- The flashing green light indicates that the device is normal state and ready to use.
- The flashing red light indicates that the device has failed the self-test and needs to maintenance. Please contact the manufacturer as soon as possible.

Check Integrity and Cleanliness

- 1) Check the integrity of the device, refer to chapter 3.2.
- 2) Check whether the device is dusty or dirty, especially the pads connector and pads connector socket.
- 3) Check whether the appearance of device have scratches or other marks of damage, especially near the pads connector and pads connector socket. If any scratches or damage are found, contact the manufacturer for maintenance.

Check Battery

In standby mode or after defibrillation, the battery power may be low.

In standby mode, the device will detect the remain battery power and validity period of the battery through self-test. If expired or low power, the status indicator flashes red and the device sends voice alarm, please review and confirmation in time.

After rescue, service personnel should check the remain battery power and the effective date. If low power or expired, replace the battery (or charge the rechargeable battery) immediately. The replaced battery should be disposed of according to local regulations. After the new battery is installed, the device automatically performs a battery installation self-test.

Caution:

After the low battery is prompted for the first time, the battery can still provide at least 30 times defibrillation. To avoid adverse impact on subsequent use, replace the battery immediately after the low battery is prompted.

Cleaning

The outside of the device can be cleaned with a soft cloth dampened in soapy water, chlorine bleach (2 tablespoons per quart or liter of water), or 96% ethanol.

Caution:

Do not immerse the device in fluids. Do not clean the accessories (Battery, Pads).

Transport

If it is necessary to transport the device to a maintenance point, the battery must be removed from the device, and packaged separately and shipped with the device. The device can be transported using general ways, but it must be protected from severe shocks, vibrations and rain and snow during transportation.

Disposal

The device and its accessories should be disposed of in accordance with local regulations.

Troubleshooting

Some common failures are listed below. They should be checked one by one in order to troubleshoot the failure. Please contact the manufacturer's designated professional personnel to repair the device.

Failure	Causes	Response	Message
Fail to turn the device on	Battery is not installed	Install the battery	N/A
	Invalid or expired battery	Replace the battery	N/A
	Mainboard error or other factors	Contact the manufacturer for maintenance	N/A
The device suddenly shut down	Invalid or expired battery	Replace battery	N/A
	Mainboard error or other factors	Contact the manufacturer for maintenance	N/A
In standby mode, the device make a Beep sound every 5 seconds, totally 5 times in continuous 25s, cycle once every hour	The device found a failure while performing self-test	Contact the manufacturer for maintenance	The device make a Beep sound every 5 seconds, totally 5 times in continuous 25s, cycle once every hour
Defibrillation charging time is too long	Battery/device failure	Stop using the device and contact manufacturer for maintenance.	<i>Charge failure</i>
	Insufficient battery	Replace battery	<i>Charge failure</i>
Voice prompt " <i>low battery</i> "	Insufficient battery	Replace battery	<i>Low battery</i>
The device cancels the charging state automatically during charging.	The pads don't stick to the patient's chest.	Attach the pads to the patient's chest	<i>Poor pads connect, check pads</i>
	Poor contact between pads and patient	Check pads contact of patient	<i>Poor pads connect, check pads</i>
	Damage of pads, cables, or pads connector	Replace pads	<i>Poor pads connect, check pads</i>
	Damage of pads socket	Contact the manufacturer for maintenance	<i>Poor pads connect, check pads</i>

Failure	Causes	Response	Message
Status indicator is switched off	Insufficient battery	Replace battery	N/A
	Damage of the status indicator	Contact the manufacturer for maintenance	The device make a Beep sound every 5 seconds, totally 5 times in continuous 25s, cycle once every hour
Data transport failed	Device failure	Contact the manufacturer for maintenance	The device make a Beep sound every 5 seconds, totally 5 times in continuous 25s, cycle once every hour.
	SIM card error	Contact the manufacturer for maintenance	N/A
Power on self-test failed	Defibrillation pads expired	Replace pads	<i>Pads overdue</i>
	Low battery/Battery expired/Battery mismatch	Replace battery	<i>Low battery/Battery Overdue/Battery Mismatch</i>
	Mainboard error or other factors	Contact manufacturer for repair	<i>Equipment failure</i>

VI Cybersecurity

This chapter mainly introduces information of cybersecurity.

Network Environment

PowerBeat M series defibrillator wakes up regularly for self-test under standby mode. After waking up, Bluetooth is turned off and 4G is turned on, and data is uploaded to the cloud server through 4G. In rescue mode and continuous VF/VT recognition mode, Bluetooth and 4G are both turned off, there is no network environment. In maintenance mode, Bluetooth and 4G are both turned on, and service personnel can config the device through Bluetooth and 4G.

- Standby mode: 4G
- Maintenance mode: Bluetooth & 4G
- Rescue mode (after the pads are connected to the patient): no network environment
- Continuous VF/VT recognition mode: no network environment

	Network Architecture	Network Type	Bandwidth
4G environment	CS	LTE-CAT1	10kbps
Bluetooth environment	CS	BLE4.0	3kbps

Data Interface

PowerBeat M series defibrillators have two external data interfaces, which are 4G and Bluetooth.

User Access

The intended use environment of PowerBeat M series defibrillators is public places or medical places, and must operated by trained professionals or emergency responders.

In addition, the management agency of the AED deployment site needs to manage and maintain the AED device to ensure that the AED can provide treatment capabilities when needed. Therefore, AED users need to be classified.

User role	Responsibility	Require	Access permission
Operator	Rescue with PowerBeat M-Series defibrillators.	Professionally trained in defibrillation and first aid	No permission
Service personnel	Install the PowerBeat M series defibrillator device, use the ViTools toolbox APP to connect to the PowerBeat M series defibrillator device, configure device parameters, upload data, and upgrade the host software.	Received professional training from the manufacturer and obtained authorization from the manufacturer	All parameters can be set

Caution:

- 1) The network interface and data interface of this device are not directly open to end users.

- | |
|--|
| <p>2) Network security related operations can only be performed by maintenance personnel or under the guidance of maintenance personnel!</p> |
|--|

Data Exchange Method

Bluetooth Transmission

PowerBeat M series defibrillator is in maintenance mode, the device turns on Bluetooth, and the mobile APP (ViTools) connects to the device via Bluetooth through authorization verification, and mainly completes the following functions through data interaction:

- 4G upload and download control
- Modify AED configuration, read and view AED configuration
- Upgrade software
- View self-test results

4G Transmission

PowerBeat M series defibrillators mainly complete the following functions through 4G under maintenance mode:

- AED configuration upload
- Resource file download and update
- Upload of rescue data

PowerBeat M series defibrillator is in standby mode and performs weekly / monthly / quarterly scheduled self-tests, it mainly completes the following functions through 4G:

- Upload of rescue data

Caution:

- 1) During the rescue data upload process, the power indicator light flashes green and red at the same time, until the rescue data transmission is completed, the power indicator light returns to the normal standby state.
- 2) If the rescue data upload fails, the device will re-upload it after the next self-test until the rescue data is uploaded successfully.

Device Security Software

No security software is required for PowerBeat M Series defibrillators.

Cyber Security Update

There are no user-required cybersecurity updates for PowerBeat M Series defibrillators.

AED Data Storage

The device will store the data during operation in the internal memory. The data type recorded by the device is as shown in the figure below:

Type of data	Data description
ECG data	ECG rhythm
Log data	Important events after the device are powered on, mainly include power on and off data, device status, rescue time, pads adhesion, button operation,

	heart rhythm analysis, charge and discharge, CPR duration, CPR operation and prompt information, discharge times and prompt information.
Self-test data	Data and results of device self-tests, including periodic self-tests, battery installation self-tests, power-on self-tests, and quick status checks.
Recording data	Audio data during rescue

ViBest-AED Data Management System

PowerBeat M series defibrillators have a built-in 4G module, which can transmit AED data to the data management system through 4G communication to achieve device management. In the data management system, service personnel can see the device status, pads validity period, battery power, location information, etc.

When the device status is abnormal or the accessories expired, the system will automatically send reminder information to service personnel for timely maintenance. After the device completes the rescue, it can upload rescue data (patient ECG data, log data, recording data, etc.) to the data management system.

Primary function	Secondary function	Describe
Front page	None	View AED maps, availability stats, and more
Basic Information	Device Information	View basic information such as device serial number, model, software version, production date, validity period, status, and accessory validity period.
Daily management	Self-test event	View AED self-inspection data on a daily, weekly, monthly, quarterly basis, and send a notification to the administrator if the self-inspection fails, such as low battery power, expired electrodes, etc.
Log management	Rescue data	Check the patient's ECG data, log data, recording data, etc. collected during the AED rescue process
System Management	User Info	View the logged in user's account, role, email, etc.
	System Message	Check the system software version number, release date, running time, etc.
	Account Management	Add, modify, and deactivate sub-accounts

Caution:

- 1) The data management system can only be operated by maintenance personnel.
- 2) The above section only describes the data management system content related to the PowerBeat M series automated external defibrillator.
- 3) For detailed instructions on the use of the ViBest-AED data management system, please contact the after-sales staff of VIVEST.

Appendix 1 Accessories

Accessories:












Name	Model	Quantity	Unit	Note
Disposable Battery (LiMnO ₂)	BAT-PT01	1	Case	Standard
Pads (PowerBeat M1)	PAZD-PT01 PAZD-PT02	1	Packet	Standard (Shipped randomly according to purchased model)
Pads (PowerBeat M3)	PAZD-PT03 PAZD-PT04	1	Packet	
Rechargeable Battery (Li-ion)	BAT-PT02	1	Case	Optional
Battery Charger	MAC01	1	Case	Optional

Packing list:















Name	Quantity	Unit
User Manual	1	Copy
Product Certification	1	Copy
Warranty Card	1	Copy
Packing List	1	Copy

Remarks: The Standard Accessories and Packing list shall be provided to the customer along with the device, and the accurate contents shall be subject to the provisions in the Packing list.

Appendix 2 Symbols

Symbol	Description
IP65	Ingress protection of the device classified as IP65 according to IEC 60529
IP54	Ingress protection of battery charger classified as IP54 according to IEC 60529
	Caution. Consult accompanying documentation
	Defibrillation-Proof Type BF Applied Part
	Warning, electricity
	General warning sign
	Operating instructions
	Follow instructions for use
	Dangerous voltage
	Stand-by
	Return to a collection site intended for Waste Electrical and Electronic Equipment (WEEE). Do not dispose of in unsorted trash
	General symbol for recovery/recyclable
	Atmospheric pressure limitation

	Temperature limitation
	Humidity limitation
	Date of manufacture
	Use by date
	This way up
	Use no hooks
	Stacking limit by number
	Fragile, handle with care
	Keep away from rain
	Do not dispose of in fire
	Do not deform or damage
	Do not re-use
	Non-sterile
	Latex free

	Batch code
	Catalogue Number
	Serial Number
	Manufacturer
	Authorized representative in the European Community
	Direct current
	Pads Icon
	Analysis Icon
	CPR Icon
	<p>Function Button:</p>  Adjusting device volume  Switching Device Modes
	Child Button
	Shock Button.



Power Button

Appendix 3 Glossaries

Glossary	Description
Standby Mode	The device will turn to standby mode after the battery installed.
Rescue Mode	The device will turn to rescue mode when the power button was pressed.
Continuous VF/VT Recognition mode	That is, the continuous detection mode, which must be operated by specialized personnel who have received training and qualifications from VIVEST to provide continuous heart rhythm detection for patients who are responsive, breathing, and have a pulse but may have cardiac discomfort. This mode only recognizes shockable and non-shockable rhythms (PowerBeat M1 only).
Pads	Contains defibrillation electrode, cable, and cable connector.
Self-test	The device uses the internal program to perform self-test on the environment where the device is located and each module in the system.
Defibrillation	The method of striking the heart with a certain current to stop ventricular fibrillation.
Pacemaker	An implantable cardiac pacing maker that stimulates the heart with electrical pulses.
Periodic self-test	When the device is in the standby mode, daily self-test, weekly self-test, and monthly self-test are performed automatically to detect batteries, internal circuits, buttons, software, etc.
Cardiac arrest	Ventricular fibrillation is the most common cause of sudden cardiac arrest due to sudden termination of ejection function.
Impedance	The device detected the electrical impedance between two pads attached to the patient's skin.
Shockable rhythm	Pulseless ventricular tachycardia or ventricular fibrillation, which can lead to cardiac arrest.
Non-shockable rhythm	The cardiac rhythm identified by the device as unsuitable for electric shock.
Sensitivity	True positive, that is, the probability that the test is not missed.
Specificity	True negative, that is, the detection of the probability of no miscarriage of justice.
Motion artifacts	Noise caused by muscle movement, cardiopulmonary resuscitation, or static electricity may interfere with cardiac analysis.

New battery	Battery that is well packed, unsealed, and valid.
Manufacturer	Unless otherwise specified, the company described in this manual is VIVEST.
ECG	Electrocardiograph.
CPR	Cardiopulmonary resuscitation, a technique for rescuing patients in cardiac arrest with artificial respiration and chest compressions.
ICD	Implantable cardioverter defibrillator
bpm	Beat per minute
AED	Automated external defibrillator
EMC	Electromagnetic Compatibility
LED	Light emitting diode
AHA	American Heart Association
SCA	Sudden Cardiac Arrest
AAMI	Association for Advancement of Medical Instrument
VF	ventricular fibrillation
VT	ventricular tachycardia

Appendix 4 Specifications

Safety Specification Features	
Safety Classification	Internally powered ME equipment
Protection against electric shock	Defibrillation-Proof Type BF Applied Part.
Particle and Water Ingress	IP65
Operational mode	Continuous operation
Security Degree	Not AP type or APG type device
ME Equipment Type	Portable
Physical parameters	
Size (height*width*length)	2.09in*3.39in*5.91in 5.3cm*8.6cm*15cm
Weight (including battery)	1.54 lb. (0.7kg)
Tolerable impact / falling damage	Free to fall from a height of 1.5 m on a hard surface
Environmental parameters	
Operation condition	Temp: -15°C to 50°C (Can work for at least 60 minutes after moving from room temperature to -20°C) Relative humidity: 0% to 95%, non-condensing Atmospheric pressure range: 50.4kPa to 106kPa
Storage condition	Short-term Temp: -40°C to 70°C (Within one week) Long-term Temp: 0°C to 50°C Relative humidity: 0% to 95%, non-condensing Atmospheric pressure range: 59.4kPa to 106kPa
Transportation condition	Short-term Temp: -40°C to +70°C (Within one week) Long-term Temp: 5°C to 50°C Relative humidity: 0% to 95%, non-condensing Atmospheric pressure range: 59.4kPa to 106kPa
Altitude	4500m
Defibrillation	
Shockable Rhythm	VT/VF
Waveform	Truncated biphasic exponential waveform


Energy level	Automatic pre-programmed selection (Adult mode: 150J; Pediatric mode: 50J)			
Output control	Manual operation (In rescue mode, the shock button should be pressed manually).			
Operational impedance limitation of patient	20Ω to 180Ω			
Charging time (Disposable battery)	Battery Status (20 ± 2°C)	from initially switching power on to readiness for maximum energy discharge	from the initial analysis to readiness for maximum discharge	from the second analysis to readiness for maximum discharge
	New Battery	Less than 19 seconds	Less than 17 seconds	Less than 8 seconds
	New battery depletes of 6 maximum energy discharge	Less than 19 seconds	Less than 17 seconds	Less than 8 seconds
	New battery depletes of 15 maximum energy discharge	Less than 19 seconds	Less than 17 seconds	Less than 8 seconds
Charging time (Rechargeable battery)	Battery Status (20 ± 2°C)	from initially switching power on to readiness for maximum energy discharge	from the initial analysis to readiness for maximum discharge	from the second analysis to readiness for maximum discharge
	New, fully charged Battery	Less than 12 seconds	Less than 10 seconds	Less than 8 seconds
	New, fully charged battery deplete of 6 maximum energy discharge	Less than 12 seconds	Less than 10 seconds	Less than 8 seconds
	New, fully charged battery deplete of 15 maximum energy discharge	Less than 12 seconds	Less than 10 seconds	Less than 8 seconds

Analysis accuracy	Comply with IEC60601-2-4 requirements	
Cardiac arrest threshold	<0.2mV	
Artifacts detection	<p>Support</p> <p>If an interfering signal that affects the accuracy of the heart rhythm analysis is detected, the device will delay performing the analysis and give a prompt. If the interference lasts for more than 20 seconds, the device will stop prompting and force the decision result.</p>	
Pads		
Specification	Combined by pads connector, wires, and a pair of disposable defibrillation electrodes.	
Length	≥1.0m	
Service life	4 years	
Battery		
Type	Disposable battery (LiMnO ₂ Battery), 12V/1500mAh (Standard) Rechargeable battery (Li-ion Battery), 7V/3450mAh (Optional)	
The number of maximum energy discharges which are available from a new and fully charged battery	<p>Disposable battery:</p> <p>a) New battery can charge and discharge at least 130 times in rated energy of 150J at 20°C±2°C environment.</p> <p>b) New battery can charge and discharge at least 20 times in rated energy of 150J at -15°C environment.</p> <p>c) new battery, after the first reminder of low battery, can charge and discharge at least 30 times.</p> <p>Rechargeable battery:</p> <p>d) New fully charged battery can charge and discharge at least 250 times in rated energy of 150J at 20°C±2°C environment.</p> <p>e) New fully charged battery can charge and discharge at least 20 times in rated energy of 150J at -15°C environment.</p> <p>f) New fully charged battery, after the first reminder of low battery, can charge and discharge at least 30 times.</p>	
standby life(Ambient temperature 20°C ± 2°C, new battery, weekly self-test, loaded into the device do not turn on the use, do not connect 4G to send self-test results.)	Disposable battery	Rechargeable battery
	3 years	3 months before recharge
Service life	3.5 years	5 years (about 300 times of charge-discharge cycle)
Battery Charger (Optional)		
Safety Classification	Class I	
Protection against	IP54	

harmful ingress of water or particulate matter	
Power Supply	AC 220V, 50Hz
Rated input power	60VA
Charge time	At 20°C ± 2°C environment, the battery charger charges 2 rechargeable batteries at the same time, and the fully charged time is not more than 3 hours and 40 minutes. A single rechargeable battery should take no more than 2 hours and 30 minutes to fully charged.
Service Life	10 years
Device	
Production date	See label on back of device
Service Life	10 years
Record	
Recording function	Immediately after the device is powered on, use the microphone to collect and record ambient sound information
Data Storage	
ECG Data	24 hours of ECG data
Self-test Data	3650 copies
Communication	
Capacity	Support mobile communication 4G and GPS positioning

Appendix 5 Defibrillation Waveform

The defibrillation waveform of the device is a truncated biphasic exponential waveform, Waveform parameters are automatically adjusted as a function of patient defibrillation impedance. In the diagram below, D is the duration of phase 1 and E is the duration of phase 2 of the waveform, F is the interphase delay (500 μ s), and P1 is the peak current.



P1: phase 1 peak current

D: Phase 1 pulse width

E: phase 2 pulse width

F: Time interval between Phase 1 and Phase 2

The device delivers shocks to load impedances from 25 to 180 ohms. The duration of each phase of the waveform is dynamically adjusted based on delivered charge, in order to compensate for patient impedance variations, as shown below:

Energy output under various impedances (Adult mode):

Load impedance (Ω)	Phase 1 pulse width D(MS) $\pm 15\%$	Phase 2 pulse width E(MS) $\pm 15\%$	Time interval between Phase 1 and Phase 2 F(MS) $\pm 15\%$	Peak current P1 (A) $\pm 15\%$	Energy output (J) $\pm 15\%$
25	2.8	2.8	0.5	61.0	128
50	4.5	4.5	0.5	33.5	150
75	6.3	5.0	0.5	23.4	155
100	8	5.3	0.5	18.0	157
125	9.7	6.4	0.5	14.5	158
150	11.5	7.7	0.5	12.0	160
175	12	8	0.5	10.5	158

Energy output under various impedances (Child mode):

Load impedance (Ω)	Phase 1 pulse width D(MS) $\pm 15\%$	Phase 2 pulse width E(MS) $\pm 15\%$	Time interval between Phase 1 and Phase 2 F(MS) $\pm 15\%$	Peak current P1 (A) $\pm 15\%$	Energy output (J) $\pm 15\%$
-----------------------------	---	---	---	-----------------------------------	---------------------------------

25	2.8	2.8	0.5	36.0	43.4
50	4.5	4.5	0.5	19.6	50.0
75	6.3	5.0	0.5	13.5	52.0
100	8.0	5.3	0.5	10.3	52.2
125	9.0	6.0	0.5	8.4	52.3
150	9.0	6.0	0.5	7.0	50.0
175	9.0	6.0	0.5	6.0	49.0

Appendix 6 ECG Analysis System

Summarize

The defibrillator's ECG analysis system, which automatically identifies patient's heart rhythm and provide shock advise to the operator, offers untrained operators the possibility of life-saving treatment in treating patients with nausea and arrhythmias. The analysis system has the following functions:

- Judgment of electrode contact
- Recognition and erasure of pacemaker signal
- Recognition of the shockable heart rhythm
- Detection of cardiac arrest
- Detection of Interference

Judgment of electrode contact

The defibrillator will automatically detect the thoracic impedance of the patient. If the impedance value is within the set threshold value, the electrode will be judged to be firmly in contact and the heart rhythm analysis can be started. If the chest impedance value exceeds the set threshold, the electrode is judged to have inadequate contact or to be improperly connected to the defibrillator, at which point the operator is advised to re-insert the electrode.

Recognition and erasure of pacemaker signal

The pulse signal of a buried pacemaker may interfere with the correct identification of arrhythmias. The defibrillator will first identify and erase the pacing signal, and then enter the rhythm analysis. Based on the results of the analysis, the shock or no shock prompt is given.

Recognition of the shockable heart rhythm

The heart rate analysis was based on the sliding ECG data window of 4S, and the analysis required about 10s. During this period, at least 4 sliding Windows were judged to be ventricular fibrillation before the final shock decision could be made, otherwise it was deemed to be no shock.

The following table illustrates the algorithmic performance of the defibrillator ECG analysis system.

Rhythm	Samples	Meet the requirements of IEC 60601-2-4 and AHA		
		Performance requirement	Actual Performance	90% one-sided lower confidence limit
Shockable rhythm – Coarse VF	231	sensitivity >90%	99.5%	99.4%
Shockable rhythm – VT	81	sensitivity >75%	98.7%	98.5%
Non-shockable rhythm – Normal Sinus Rhythm	362	specificity >99%	100%	100%
Non-shockable rhythm – Asystole	103	specificity >95%	100%	100%

Non-shockable rhythm – all other non-shockable rhythm	520	specificity >95%	99.1%	99.1%
---	-----	------------------	-------	-------

Data Source: PhysioNet, a public database of adult ECG data collected from various hospitals.


Detection of cardiac arrest

Pause threshold is peak-peak 0.2mV. Be careful that the electrical signal peak pea is less than 0.2mV, the system will recognize the pause and give a warning that electric shock is not recommended, and CPR will be initiated.

Detection of Interference

The defibrillator's ECG analysis system detects interference, which may be caused by external sources such as posture movements or electrical noise. Postural movement includes patient movement, rescuer movement, vehicle movement, etc.; External sources of electronic noise: e.g., mobile phones, radios, etc. If interference is detected, the system sends a voice warning to the rescuer, at which point the operator should remove the interference as soon as possible to minimize artifacts in the ECG, and the system continues to perform heart rate analysis.

Appendix 7 EMC Guide

 Warning	<ol style="list-style-type: none"> 1) Except for accessories sold by the manufacturer of the device, using accessories that are not specified may result in an increase in device emission or a decrease in immunity. 2) The device should not be used in proximity to or stacked with other devices. If it must be used in proximity to or stacked with other devices, it should be observed to verify that it works properly in its configuration. 3) You need to protect the device EMC and install and maintain it in an environment that meets the following EMC information. 4) Even if other equipment meets CISPR's launch requirements, it can cause interference with the device. 5) Other devices that contain RF radiofrequency emission may affect this device (for example, mobile phones with wireless computers).
--	--

ESSENTIAL PERFORMANCE:

The essential performance of PowerBeat M series AED is to deliver defibrillation therapy and accurately differentiate between shockable and non-shockable rhythms.

Electromagnetic Emissions


Guidelines and manufacturer's claims for Electromagnetic Launches		
The device is intended for use in the following specified electromagnetic environments, and the purchaser or user should ensure that it is used in this electromagnetic environment:		
Emission Test	COMPLIANCE	ELECTROMAGNETIC ENVIRONMENT- GUIDANCE
Radio-frequency emission CISPR 11	Group 1	The device uses RF energy only for its internal functions, so its RF emission is low and the likelihood of interference with nearby electronics is minimal.
Radio-frequency emission CISPR 11	Class B	The device is suitable for use in all facilities, including domestic facilities and directly connected to the domestic residential public low voltage power grid.

Electromagnetic Immunity

Guidelines and manufacturer's declaration of Electromagnetic Immunity			
The device is intended for use in the following specified electromagnetic environments, and the purchaser or user should ensure that it is used in this electromagnetic environment:			
IMMUNITY TEST	IEC 60601 TEST LEVEL	COMPLIANCE LEVEL	ELECTROMAGNETIC ENVIRONMENT - GUIDANCE
Electrostatic Discharge IEC 61000-4-2	$\pm 6\text{kV}$ contact $\pm 8\text{kV}$ air	$\pm 6\text{kV}$ contact $\pm 8\text{kV}$ air	Floors should be wood, concrete or tile. The relative humidity should be at least 30% if they are covered with synthetic materials.
PFMF (50Hz/60Hz) IEC 61000-4-8	3A/m	3A/m	The power-frequency magnetic field shall have the horizontal characteristics of a typical site in a typical commercial or hospital environment.
Note: U_T refers to the AC network voltage before the test voltage is applied.			

Guidelines and manufacturer's declaration of Electromagnetic Immunity

The device is intended for use in the following specified electromagnetic environments, and the purchaser or user should ensure that it is used in this electromagnetic environment:

IMMUNITY TEST	IEC 60601 TEST LEVEL	COMPLIANCE LEVEL	ELECTROMAGNETIC ENVIRONMENT - GUIDANCE
<p>Conducted RF IEC 61000-4-3</p> <p>Radiated RF IEC 61000-4-3</p>	<p>3V (effective value) 150kHz~80MHz (Expect ISM bands ^a)</p> <p>10V (effective value) 150kHz~80MHz (ISM bands ^a)</p> <p>10V/m 80MHz~2.5GHz</p>	<p>3V (effective value)</p> <p>10V (effective value)</p> <p>10V/m</p>	<p>Portable and mobile RF communications equipment should not be used closer to any part of the device, including cables, than the recommended isolation distance, which is calculated by a formula corresponding to the transmitter frequency.</p> <p>Recommended separation distance:</p> <p>$d=1.17 \sqrt{P}$</p> <p>$d=1.20 \sqrt{P}$</p> <p>$d=1.2\sqrt{P}$ 80MHz~800MHz $d=2.3\sqrt{P}$ 800MHz~2.5GHz</p> <p>In which:</p> <p>P——Based on the transmitter manufacturer's maximum transmitter output rating in watts (W).</p> <p>d——Is the recommended isolation distance in meters (m) ^b.</p> <p>The field intensity of a stationary RF transmitter is determined by surveying the electromagnetic field ^c and should be lower than the coincidence level in each frequency range^d.</p> <p>Interference may occur near equipment marked with the following symbols: </p>

Note:

- 1) In 80 MHz and 800 MHz environment, the formula of higher frequency band is adopted.
- 2) These guidelines are not suitable for all situations. Electromagnetic propagation is affected by

absorption and reflection from building objects and human bodies.

a: The engineering medical frequency band between 150kHz and 80MHz refers to 6.765MHz~6.795MHz 13.553MHz~13.567MHz 26.957MHz~27.283MHz and 40.66MHz~40.70MHz.

b: Coincidence levels in the engineering medical frequency band between 150kHz and 80MHz and in the frequency range of 80MHz to 2.5GHz are used to reduce the possibility of interference caused by mobile/portable communication devices being accidentally brought into the patient area. For this reason, the additional factor 10/3 is used to calculate the recommended isolation distance for transmitters in these frequency ranges

c: Stationary transmitter, such as: wireless cellular/cordless phones and ground mobile radio base station Am and FM radio amateur radio and television broadcasting, etc., could not accurately predict the field intensity in theory, to assess stationary rf transmitter electromagnetic environment, electromagnetic site survey should be taken into account. If the measured:

If the field intensity of PowerBeat M series AED is higher than the applicable RF compliance level above, the device should be observed to verify its normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorientation or positioning of the device.

d: In the whole frequency range of 150kHz~80MHz, the field intensity should be lower than 3V/m

Recommended isolation distances between portable and mobile RF communications equipment and PowerBeat M series AED

The device is intended for use in electromagnetic environments where rf disturbance is controlled. Depending on the maximum rated output power of the communication equipment, the purchaser or user can protect against EMI by maintaining the minimum distance between portable and mobile RF communication equipment (transmitters) and the device as recommended below:

Rated maximum output power of transmitter(W)	Separation distance according to frequency of transmitter(m)			
	150kHz~80MHz (Except ISM band) $d=1.17\sqrt{P}$	150kHz~80MHz (ISM band) $d=1.2\sqrt{P}$	80MHz~800MHz $d=1.2\sqrt{P}$	800MHz~2.5GHz $d=2.3\sqrt{P}$
0.01	0.12	0.12	0.12	0.23
0.1	0.37	0.38	0.38	0.73
1	1.17	1.20	1.20	2.3
10	3.70	3.79	3.79	7.27
100	11.70	12.00	12.00	23.00

For the maximum rated output power of the transmitter not listed in the above table, the recommended isolation distance D, in meters (m), can be determined by the formula in the corresponding transmitter frequency column, where P is the maximum rated output power of the transmitter provided by the transmitter manufacturer, in watts (W).

Note:

- 1) At 80MHz and 800MHz frequency points, the formula of higher frequency band is adopted
- 2) The engineering medical frequency band between 150kHz and 80MHz refers to 6.765 MHz~6.795MHz 13.553MHz~13.567MHz 26.957MHz~27.283MHz and

40.66MHz~40.70MHz

- 3) The additional factor 10/3 is used to calculate the recommended isolation distance for transmitters in the 150kHz to 80MHz engineering medical band and 80MHz to 2.5GHz frequency range to reduce the possibility of interference caused by portable/mobile communication equipment being accidentally brought into the patient area
- 4) These guidelines may not be appropriate in all cases where electromagnetic propagation is affected by absorption and reflection from building objects and human bodies

Regulatory Compliance

VIVEST solemnly declares that PowerBeat M series Automated External Defibrillator complies with the relevant provisions of the relevant medical equipment standards:

IEC 60601-1:2005+A1:2012 Medical electrical equipment – Part 1: General requirements for basic safety and essential performance,

IEC 60601-2-4:2010+AMD1:2018 Medical electrical equipment – Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators,

IEC 60601-1-2:2014 Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests,

IEC 60601-1-12:2014 Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment.